



Correction Table  
for  
Hair operated instruments

CASELLA  
LONDON

ADD  
this correction algebraically  
to Relative Humidity reading

°C	↓	°F
20.0	0	68.0
22.5	1	72.5
25.0	2	77.0
27.5	3	81.5
30.0	4	86.0
32.5	5	90.5
35.0	6	95.0
37.5	7	99.5
40.0	8	104.0
42.5	9	108.5
45.0	10	113.0
47.5	11	117.5
50.0	12	122.0